

# Test Recall iFOB



## Clearview iFOB Complete Test Recall

The manufacturer of the “Clearview iFOB Complete” is conducting a recall of certain lots of collection and testing devices. These devices are being recalled due to reduced sensitivity of the test to detect blood in feces, potentially causing false negative results.

Until this issue is resolved, St. Luke’s Laboratory will be using the Hemocult SENSE method of testing for blood in feces. Kits are being supplied to all of our clinics and outpatient areas to replace the iFOB kits that are currently in these locations.

St. Luke’s Laboratory will notify physicians if they have any patients who have been affected by this recall. These patients should be retested with the Hemocult SENSE method. This retesting will be done at no charge.

**If you have questions, please contact K. Baer, M.D. (218)249-5751, Chemistry Medical Director or Sue Bachinski , MT(ASCP) (218) 249-2445 Laboratory Outreach Manager.**